



JWE GmbH

Dental Floss Abrasion Test

Function

The floss abrasion tester allows comparative study of dental floss materials for strength and abrasion at constant experimental conditions in the laboratory. The floss is moved with a specific load several times on a polishing paper. The polishing paper is mounted on an idealized tooth.

The device includes the cleaning movement mechanism with the fastening of the floss and an idealized molar tooth with polishing paper and clamping .

On the front are the display and controls. The speed and the number of cycles can be adjusted.



Specification

Dimension (W,D,H) ca.	400, 300, 400 mm
Weight	50 kg
Cycle Speed	1-45 Zyklen/min
Cycles	1 – 1000 n
Stroke	45 mm
Load	1 – 5.5 N
Test Sample	10 Stk.
Power Supply	115/230 (50/60 Hz)

Ordering Information

Device	309-010
--------	---------

Scope of Delivery

- 1 Device
- 1 Operating Instruction
- 1 Power Cord CEE 16A (5 m)

Conditions of Delivery

Delivery:	EXW
Shipping time:	by arrangement

Contact

JWE GmbH
Tel. 0049 (0)6173 3948596
Fax 032 22 1455 209
Email info@jw-e.net
Web www.jw-e.net

Technical changes reserved.
The illustrations may differ from the actual equipment.
All data without obligation and not-binding.